

PRESSURE TRANSDUCER  
FINAL CALIBRATION0.00 - 30.00 PSIA  
Excitation 28.000 VdcJob: TN1130  
Model: PX309-030AI  
Date: 1/28/2010

Serial: 062008I058

Tested By: JP

Temperature Range: 4 to 185 F.

Calibrated: 0.00 - 30.00 PSIA

Specfile: PX309-I+=15-=50G-=30

Pressure PSIA	Unit Data mA
0.00	4.0140
15.00	12.0209
30.00	20.0067
15.00	12.0305
0.00	4.0148

Balance	4.0140	mA
Sensitivity	15.9927	mA

ELECTRICAL LEAKAGE: PASS

PRESSURE CONNECTION/FITTING: 1/4-18 NPT Male

ELECTRICAL WIRING/CONNECTOR: BLACK = -EXC  
RED = +EXC

This Calibration was performed using Instruments and Standards that are traceable to the United States National Institute of Standards Technology.

S/N	Description	Range	Reference	Cal Cert
1742/94-8	AUTO 150 PSI DRUCK	0 - 150 PSIA	C-2509	0906251742
MY44026132	AT34970A DMM UUT	Unit Under Test	C-1310	908949730

Q.A. Representative : John Piendl

Date: 1/28/2010

This transducer is tested to & meets published specifications. After final calibration our products are stored in a controlled stock room & considered in bonded storage. Depending on environment & severity of use factory calibration is recommended every one to three years after initial service installation date.

Omegadyne, Inc., 149 Stelzer Court, Sunbury, OH 43074 (740) 965-9340  
<http://www.omegadyne.com> email: [info@omegadyne.com](mailto:info@omegadyne.com) (800) USA-DYNE